

FA-2 N3 Fix Replenisher

For Processing C-41 and Compatible Color Negative Films Through Process C-41SM



Catalog Number: 600005420 (4 x 300 Rolls 135-24)

Approximate based upon 1.118 meters/roll of 135-24 film

DESCRIPTION

This product is designed for use in the FA-2 System N3 Fixer reservoir tank installed in any Noritsu C-41SM minilab processor.

SAFE HANDLING INSTRUCTIONS

Please refer to Material Safety Data Sheets (MSDS) for specific chemical details. For emergencies, please contact CHEMTREC at 1-800-424-9300. You can obtain an MSDS by calling the Fujifilm Technical Hotline at 1-800-526-0851 Ext. 10.

HMIS INFORMATION

Hazardous Materials Identification System. Please refer to Section 3 in the MSDS for further information.

	Concentrate	Working Tank
	Part A	
Health	1	1
Flammability	0	0
Reactivity	0	0
Personal Protection	C	C

SPECS & RECOMMENDATIONS

Temperature	35 - 41° C 95 - 106° F	
Processing Time	2'00"	
Starting Replenishment Rate	135-24	
	15.3 ml/m	
	STARTING PUMP SETTING	
	Fix	Fix-W
	4.50	10.80

SPECIFIC GRAVITY & pH

	pH @ 25° C	SpG @ 20° C	SpG @ 37.8° C
Replenisher Tank	Concentrates – Specifications Not Published		
Fresh Tank	7.40 ± 0.20	1.123 ± 0.020	1.115 ± 0.020
Seasoned Tank #1	5.90 ± 0.20	1.110 ± 0.015	1.102 ± 0.015
Seasoned Tank #2	6.70 ± 0.20	1.107 ± 0.015	1.098 ± 0.015

INSTALLATION INSTRUCTIONS

N3 (Fix) Fixer Replenisher



- Determine which FA2 System reservoir is empty by identifying which light is flashing red.
- Remove the cap – **not the foil seal** – from the corresponding chemical concentrate bottle.
- Open the corresponding FA2 System hinged lid to reveal the foil puncture knife.
- Turn the chemical concentrate bottle upside down, position the bottle so it slides into the bottle opening easily and push the bottle down onto the foil puncture knife.
- Allow the concentrate in the bottle to completely drain into the FA2 reservoir.
- After a few seconds the flashing light will turn solid red, then solid green, indicating the reservoir is full.
- Carefully remove the empty bottle and dispose of properly. Reset the drawer by turning the front knob **Off**, then **On** again, before closing the drawer.

SHELF LIFE

	Concentrate	Working Tank
Time	U = 24 months O = 2 weeks	2 Weeks
Color	Clear, Pale Yellow	Clear, Pale Yellow
Odor	Slight Vinegar Odor	Slight Ammonia Odor

U = Unopened, O = Opened

The data contained in this photographic bulletin is believed to be true and accurate, but is offered solely for your consideration, investigation, and verification. Nothing herein shall be construed to be a warranty or guarantee by either FUJIFILM Hunt Chemicals or any of its affiliates, and all such warranties, implied or otherwise, including any implied warranty of merchantability, are hereby expressly disclaimed. You, of course, are fully responsible for any use and / or domestic or foreign sale of the product(s) described, and nothing in this photographic bulletin shall be construed to constitute permission or a recommendation to practice any invention covered by patent or patent application or know-how owned by FUJI HUNT, its affiliates, or others.

November 2009

Fujifilm Hunt Chemicals U.S.A., Inc.
40 Boroline Road
Allendale, New Jersey 07401

FUJIFILM

Technical Hotline
1-800-526-0851 Ext. 10